

INSTRUCTION MANUAL



Model HFP01 Soil Heat Flux Plate

12/09



Copyright © 2002-2009
Campbell Scientific, Inc.

WARRANTY AND ASSISTANCE

This equipment is warranted by CAMPBELL SCIENTIFIC (CANADA) CORP. ("CSC") to be free from defects in materials and workmanship under normal use and service for **twelve (12) months** from date of shipment unless specified otherwise. ***** **Batteries are not warranted.** ***** CSC's obligation under this warranty is limited to repairing or replacing (at CSC's option) defective products. The customer shall assume all costs of removing, reinstalling, and shipping defective products to CSC. CSC will return such products by surface carrier prepaid. This warranty shall not apply to any CSC products which have been subjected to modification, misuse, neglect, accidents of nature, or shipping damage. This warranty is in lieu of all other warranties, expressed or implied, including warranties of merchantability or fitness for a particular purpose. CSC is not liable for special, indirect, incidental, or consequential damages.

Products may not be returned without prior authorization. To obtain a Return Merchandise Authorization (RMA), contact CAMPBELL SCIENTIFIC (CANADA) CORP., at (780) 454-2505. An RMA number will be issued in order to facilitate Repair Personnel in identifying an instrument upon arrival. Please write this number clearly on the outside of the shipping container. Include description of symptoms and all pertinent details.

CAMPBELL SCIENTIFIC (CANADA) CORP. does not accept collect calls.

Non-warranty products returned for repair should be accompanied by a purchase order to cover repair costs.



CAMPBELL SCIENTIFIC
C A N A D A C O R P .

11564 - 149 street - edmonton - alberta - T5M 1W7
tel 780.454.2505 fax 780.454.2655

www.campbellsci.ca

PLEASE READ FIRST

About this manual

Please note that this manual was originally produced by Campbell Scientific Inc. (CSI) primarily for the US market. Some spellings, weights and measures may reflect this origin.

Some useful conversion factors:

Area: 1 in² (square inch) = 645 mm²

Length: 1 in. (inch) = 25.4 mm
1 ft (foot) = 304.8 mm
1 yard = 0.914 m
1 mile = 1.609 km

Mass: 1 oz. (ounce) = 28.35 g
1 lb (pound weight) = 0.454 kg

Pressure: 1 psi (lb/in²) = 68.95 mb

Volume: 1 US gallon = 3.785 litres

In addition, part ordering numbers may vary. For example, the CABLE5CBL is a CSI part number and known as a FIN5COND at Campbell Scientific Canada (CSC). CSC Technical Support will be pleased to assist with any questions.

HFP01 Table of Contents

PDF viewers note: These page numbers refer to the printed version of this document. Use the Adobe Acrobat® bookmarks tab for links to specific sections.

1. General Description	1
2. Specifications	1
3. Installation	2
4. Wiring	3
5. Example Programs	4
6. Soil Heat Flux and Storage	6
7. Maintenance	7
8. References	7

Appendix

A. General Theory of Heat Flux Sensors	A-1
A.1 General Theory	A-1
A.2 Extended Theory.....	A-2

Figures

1. Placement of Heat Flux Plates	2
2. HFP01 Plate to Datalogger Connections	3
A-1. General Characteristics of a Heat Flux Sensor	A-1
A-2. Deflection Error.....	A-2

Tables

1. Datalogger Connections for a Single-Ended Measurement.....	3
2. Wiring for Example 1	4
3. Wiring for Example 2	5
4. Wiring for Example 3	6

Model HFP01 Soil Heat Flux Plate

1. General Description

The HFP01 Soil Heat Flux Plate uses a thermopile to measure temperature gradients across its plate. Operating in a completely passive way, it generates a small output voltage that is proportional to this differential temperature. Assuming that the heat flux is steady, that the thermal conductivity of the body is constant, and that the sensor has negligible influence on the thermal flow pattern, the signal of the HFP01 is directly proportional to the local heat flux.

The HFP01's output is in millivolts. To convert this measured voltage to heat flux, it must be divided by the plate's calibration constant. A unique calibration constant is supplied with each sensor.

To get the soil heat flux at the surface, use at least two HFP01s to measure soil heat flux at a depth of 8 cm; a TCAV Averaging Soil Thermocouple to measure the temporal change in temperature of the soil layer above the HFP01; and a CS616 Water Content Reflectometer to measure soil water content (see Figure 1). The temporal change in soil temperature and soil water content are used to compute the soil storage term.

The -L extension in the model name (i.e., HFP01-L) indicates a user-specified cable length. This manual refers to the sensor as the HFP01.

NOTE

A general discussion about the characteristics and principles of operation of heat flux sensors is shown in Appendix A.

2. Specifications

Operating Temperature:	-30°C to +70°C
Plate Thickness:	5.0 mm (0.20 in)
Plate Diameter:	80.0 mm (3.15 in)
Weight (w/o cable):	7.05 oz (200 g)
Sensor:	Thermopile
Measurement Range:	$\pm 2000 \text{ W m}^{-2}$
Sensitivity (nominal):	$50 \mu\text{V W}^{-1} \text{ m}^{-2}$
Expected Typical Accuracy (12 hour totals):	within -15% to +5% in most common soils
Nominal Resistance:	2 W
Sensor Thermal Resistance:	$< 6.25 \times 10^{-3} \text{ Km}^2\text{W}^{-1}$




FIGURE 1. Placement of Heat Flux Plates

3. Installation

The standard set of sensors for measuring soil heat flux includes two HF01 Soil Heat Flux Plates, one TCAV Averaging Soil Thermocouple, and one CS616 Water Content Reflectometer. These sensors are installed as shown in Figure 1.

The location of the heat flux plates and thermocouple should be chosen to be representative of the area under study. If the ground cover is extremely varied, an additional set of sensors is required to provide a valid soil heat flux average.

Use a small shovel to make a vertical slice in the soil. Excavate the soil to one side of the slice. Keep this soil intact so that it can be replaced with minimal disruption.

The sensors are installed in the undisturbed face of the hole. Measure the sensor depths from the top of the hole. With a small knife, make a horizontal cut 8 cm below the surface into the undisturbed face of the hole. Insert the heat flux plate into the horizontal cut.

NOTE

Install the HFP01 in the soil such that the side with the red label is facing the sky and side with a blue label facing the soil.

CAUTION

In order for the HFP01 to make quality soil heat flux measurements, the plate must be in full contact with the soil.

Never run the sensors leads directly to the surface. Rather, bury the sensor leads a short distance back from the hole to minimize thermal conduction on the lead wire. Replace the excavated soil back into its original position after all the sensors are installed.

NOTE To protect sensor cables from damage caused by rodents, it is recommended to bury them inside of flexible electrical tubing.




FIGURE 2. HFP01 Plate to Datalogger Connections

4. Wiring

Connections to Campbell Scientific dataloggers are given in Table 1. The output of the HF01 can be measured using a single-ended analog measurement (VoltSE in CRBasic or Instruction 1 in Edlog) or a differential analog measurement (VoltDiff in CRBasic or Instruction 2 in Edlog).

Description	Color	CR10(X), CR510, CR500	CR800, CR850, CR1000, CR3000, CR5000, CR23X, 21X, CR7,
Signal	White	Single-Ended Input	Single-Ended Input
Signal Reference	Green	AG	⊕
Shield	Clear	G	⊕

5. Example Programs

To calculate the calibration multiplier, divide 1000 by the nominal calibration sensitivity (i.e., 1000/sensitivity). The nominal calibration sensitivity is unique for each HFP01. Each sensor is accompanied with a calibration certificate and has a label on it with the calibration values. The label is located on the pigtail end of the sensor leads.

In example 1, the nominal calibration sensitivity is 67.1 $\mu\text{V}/\text{W}^{-2}$ for the HFP01#1 and 67.0 $\mu\text{V}/\text{W}^{-2}$ for the HFP01#2. In examples 2 and 3, the nominal calibration sensitivity is 67.1 $\mu\text{V}/\text{W}^{-2}$.

Description	Color	CR3000
Sensor Signal #1	White	7H
Sensor Signal Reference #1	Green	7L
Shield #1	Clear	⚡
Sensor Signal #2	White	8H
Sensor Signal Reference #2	Green	8L
Shield #2	Clear	⚡

Example 1. Sample CR3000 Program using Differential Measurement Instructions

```
'CR3000 Series Datalogger
'Copyright (c) 2007, Campbell Scientific, Inc. All rights reserved.
'This datalogger program measures two HFP01 Hukseflux soil heat flux sensors
*** Wiring ***
'7H    HFP01 #1 signal (white)
'7L    HFP01 #1 signal reference (green)
'gnd   HFP01 #1 shield (clear)
'8H    HFP01 #2 signal (white)
'8L    HFP01 #2 signal reference (green)
'gnd   HFP01 #2 shield (clear)
*** Constants ***
Const HFP01_CAL_1 = 14.90    'Unique multiplier for HFP01 #1. 1000/sensitivity (1000/67.1)
Const HFP01_CAL_2 = 14.92    'Unique multiplier for HFP01 #2. 1000/sensitivity (1000/67.0)
*** Variables ***
```

```
'Energy balance sensors.

Public shf(2)
Dim shf_cal(2)
Units shf = W/m^2

'*** Final Output Data Tables ***

DataTable (soil_hf,TRUE,-1)
    DataInterval (0,30,Min,10)

    Average (2,shf(1),IEEEE4,FALSE)

EndTable

'*** Program ***

BeginProg

'Load the HFP01 factory calibration.
shf_cal(1) = HFP01_CAL_1
shf_cal(2) = HFP01_CAL_2

Scan (1,Sec,3,0)

'Measure the HFP01 soil heat flux plates.
VoltDiff (shf(1),2,mV50C,7,TRUE,200,250,shf_cal(),0)

CallTable soil_hf

NextScan
EndProg
```

Description	Color	CR10(X)
Signal	White	SE 5 (3H)
Signal Reference	Green	AG
Shield	Clear	G

Example 2. Portion of CR10(X) Program using the Single-Ended Measurement Instruction

NOTE The instruction below does not store data in final storage. P92, P77, and an output processing instruction are required to store the data permanently.

```
01: Volt (SE) (P1)
1: 1 Reps
2: 2 7.5 mV Slow Range ;CR510 (7.5 mV);CR23X (10 mV); 21X, CR7 (5 mV)
3: 5 SE Channel ;White wire (SE 5), Green wire (AG)
4: 1 Loc [ HFP01 ]
5: 14.90 Mult ;Enter Calibration
6: 0 Offset
```

TABLE 4. Wiring for Example 3		
Description	Color	CR23X
Signal	Green	9H
Signal Reference	White	9L
Shield	Clear	≠

Example 3. Portion of CR23X Program using the Differential Measurement Instruction

NOTE The instruction below does not store data in final storage. P92, P77, and an output processing instruction are required to store the data permanently.

```

;Measure the HFP01 Soil Heat Flux plate.
;
01: Volt (Diff) (P2)
  1: 1      Repts
  2: 21     10 mV, 60 Hz Reject, Slow Range ;CR510, CR10(X) (7.5 mV); 21X, CR7 (5 mV)
  3: 9      DIFF Channel ;White wire (9H); Green wire (9L)
  4: 1      Loc [ HFP01 ]
  5: 14.90  Mult ;Enter Calibration
  6: 0      Offset
    
```

6. Soil Heat Flux and Storage

The soil heat flux at the surface is calculated by adding the measured flux at a fixed depth, d , to the energy stored in the layer above the heat flux plates. The specific heat of the soil and the change in soil temperature, ΔT_s , over the output interval, t , are required to calculate the stored energy.

The heat capacity of the soil is calculated by adding the specific heat of the dry soil to that of the soil water. The values used for specific heat of dry soil and water are on a mass basis. The heat capacity of the moist soil is given by:

$$C_s = \rho_b (C_d + \theta_m C_w) = \rho_b C_d + \theta_v \rho_w C_w \tag{1}$$

$$\theta_m = \frac{\rho_w}{\rho_b} \theta_v \tag{2}$$

where C_s is the heat capacity of moist soil, ρ_b is bulk density, ρ_w is the density of water, C_d is the heat capacity of a dry mineral soil, θ_m is soil water content on a mass basis, θ_v is soil water content on a volume basis, and C_w is the heat capacity of water.

This calculation requires site specific inputs for bulk density, mass basis soil water content or volume basis soil water content, and the specific heat of the

dry soil. Bulk density and mass basis soil water content can be found by sampling (Klute, 1986). The volumetric soil water content is measured by the CS616 water content reflectometer. A value of $840 \text{ J kg}^{-1} \text{ K}^{-1}$ for the heat capacity of dry soil is a reasonable value for most mineral soils (Hanks and Ashcroft, 1980).

The storage term is then given by Eq. (3) and the soil heat flux at the surface is given by Eq. (4).

$$S = \frac{\Delta T_s C_s d}{t} \quad (3)$$

$$G_{sfc} = G_{8cm} + S \quad (4)$$

7. Maintenance

The HFP01 requires minimal maintenance. Check the sensor leads monthly for rodent damage.

Recalibrate the HFP01 for every two years of continuous use. Obtain an RMA number before returning the HFP01 to Campbell Scientific for calibration.

8. References

- Hanks, R. J., and G. L. Ashcroft, 1980: *Applied Soil Physics: Soil Water and Temperature Application*. Springer-Verlag, 159 pp.
- Klute, A., 1986: *Method of Soil Analysis*. No. 9, Part 1, Sections 13 and 21, American Society of Agronomy, Inc., Soil Science Society of America, Inc.

Appendix A. General Theory of Heat Flux Sensors

This Appendix discusses the general theory and characteristics of heat flux sensors similar to the HFP01.

A.1 General Theory




FIGURE A-1. General Characteristics of a Heat Flux Sensor

When heat is flowing through the sensor in the indicated direction, the filling material will act as a thermal resistance. Consequently the heat flow, ϕ , will follow a temperature gradient across the sensor, flowing from the hot to the cold side. Most Heat Flux Sensors are based on a thermopile—a number of thermocouples connected in series. A single thermocouple will generate an output voltage that is proportional to the temperature difference between the joints (copper-constantan and constantan-copper). Provided that errors are avoided, the temperature difference is proportional to the heat flux—depending only on the thickness and the thermal conductivity of the sensor. Using more thermocouples in series will enhance the output signal. In Figure A-1, the joints of a copper-constantan thermopile are alternatively placed on the hot and the cold side of the sensor.

The thermopile is embedded in a filling material, usually a plastic.

Each individual sensor will have its own sensitivity, E_{sen} , usually expressed in Volts output, V_{sen} , per Watt per square metre heat flux, ϕ . The flux is calculated:

$$\phi = V_{sen} / E_{sen}$$

The sensitivity is determined by the manufacturer, and is found on the calibration certificate that is supplied with each sensor.

When used for measuring soil heat flux, heat flux sensors such as the HFP01 reach a limited level of accuracy. This has to do with the fact that thermal parameters of soil are constantly changing (soil moisture content) and with the fact that the ambient temperature is not fixed. A realistic estimate of the error range is $\pm 20\%$ over a thermal conductivity range from 0.1 to 1.7 W/mK (dry sand to water-saturated sand) across the temperature range of -30° to $+70^\circ\text{C}$. The accuracy is

better if the soil conditions are closer to the reference conditions (see the sensor specifications) and, in an actual experiment, the expected error range will probably be $\pm 10\%$.

The reference conditions for the calibration are a thermal conductivity of 0.8 W/mK and a nominal temperature of 20°C.

A.2 Extended Theory

It is obvious that there is the possibility that the sensor itself can significantly disturb the phenomenon that it is supposed to measure. By adding a sensor to the material under observation, you can add additional, and sometimes differing, thermal resistances.

The deflection error, as shown in Figure A-2, represents the effect that, as a result of differing resistances, the flow pattern will change, especially at the edges of the Heat Flux Sensor. The order of magnitude of this error for strongly different thermal conductivity values between the sensor and its environment (for example 0.6 for a typical sensor and 0.03 for an insulating wall) is about 40%.




FIGURE A-2. Deflection Error

The heat flux is deflected at the edges of the sensor. As a result, the heat flow at the edges is not representative.

Apart from the sensor thermal resistance, the contact resistance between the sensor and surrounding material require special attention.

The conductivity of air is approximately 0.02W/mK which is ten times smaller than that of the Heat Flux Sensor. It follows, therefore, that air gaps can form major contact resistances. In all cases the contact between sensor and surrounding material should be as close and as stable as possible, so that it will not influence the measurements.

The aspects of differing thermal properties between sensor and its environment can also be dealt with during the measurement using a higher accuracy 'Self-Calibrating' type of Heat Flux Sensor.

Campbell Scientific Companies

Campbell Scientific, Inc. (CSI)

815 West 1800 North
Logan, Utah 84321
UNITED STATES
www.campbellsci.com • info@campbellsci.com

Campbell Scientific Africa Pty. Ltd. (CSAf)

PO Box 2450
Somerset West 7129
SOUTH AFRICA
www.csafrica.co.za • cleroux@csafrica.co.za

Campbell Scientific Australia Pty. Ltd. (CSA)

PO Box 444
Thuringowa Central
QLD 4812 AUSTRALIA
www.campbellsci.com.au • info@campbellsci.com.au

Campbell Scientific do Brazil Ltda. (CSB)

Rua Luisa Crapsi Orsi, 15 Butantã
CEP: 005543-000 São Paulo SP BRAZIL
www.campbellsci.com.br • suporte@campbellsci.com.br

Campbell Scientific Canada Corp. (CSC)

11564 - 149th Street NW
Edmonton, Alberta T5M 1W7
CANADA
www.campbellsci.ca • dataloggers@campbellsci.ca

Campbell Scientific Centro Caribe S.A. (CSCC)

300 N Cementerio, Edificio Breller
Santo Domingo, Heredia 40305
COSTA RICA
www.campbellsci.cc • info@campbellsci.cc

Campbell Scientific Ltd. (CSL)

Campbell Park
80 Hathern Road
Shepshed, Loughborough LE12 9GX
UNITED KINGDOM
www.campbellsci.co.uk • sales@campbellsci.co.uk

Campbell Scientific Ltd. (France)

Miniparc du Verger - Bat. H
1, rue de Terre Neuve - Les Ulis
91967 COURTABOEUF CEDEX
FRANCE
www.campbellsci.fr • info@campbellsci.fr

Campbell Scientific Spain, S. L.

Psg. Font 14, local 8
08013 Barcelona
SPAIN
www.campbellsci.es • info@campbellsci.es

Please visit www.campbellsci.com to obtain contact information for your local US or International representative.